

	Questions	Answers
1	<p>1. Regarding your requirements for 11.2. Envelope B: "Technical offer"-</p> <ul style="list-style-type: none"> • For each lot (01, 02 ad 03) the tenderer intend to submit the offer, the description of the team (key staff is sufficient) in charge of the execution of the contract (using the specific template included in Envelope B); • For each lot (01, 02 ad 03) the tenderer intend to submit the offer, the curricula of the team members (key staff is sufficient) in charge of the execution of the contract (using the specific template included in Envelope B). If the curricula proposed do not match with the requirements contained in the Technical Specifications, the Contracting Authority reserves the right to request the tenderer new curricula; <p>Do we need to fill only template in Envelope B „technical team and CV„, to confirm this two requirements or we need to submit additional papers or information?</p>	<p>1</p> <ul style="list-style-type: none"> • The format supplied in envelop B is mandatory. Additional documentation is accepted. • The format supplied in envelop B is mandatory. Additional documentation is accepted.
2	<p>Can you please inform us in case of consortium or joint venture does every person/ company must comply with the requests in 11.1. Envelope A and 11.2. Envelope B or if one of the members its compliat to them it applies to all as consortium/joint venture?</p>	<p>As regards point 11.1 envelope A, in the event that a consortium is formed, each member of the consortium needs to satisfy the non-exclusion as well as the selection criteria indicated respectively in points 13 and 19 of the contract notice. In this regard, each member of the consortium needs to fill out the tenderer's declaration in point 6 of the tender form. Along with such declaration each member of the consortium needs to report separately the data concerning its individual financial capacity by using the table layout in point 3 of the tender form and adding it at the conclusion of the tender's declaration. Each member of the consortium needs also to annex the declaration of honour to the tender's declaration. As far as a consortium is concerned, the data reported in the table in point 3 of the tender form must be the sum of the data</p>

		reported in the corresponding tables in the declarations provided by the consortium members. Note that point 13 of the tenderer's declaration makes erroneously reference to point 9 instead of point 13 of the contract notice (exclusion criteria). While sending the tenderer's declaration, include the correct reference. As regards point 11.2 envelope B, the technical offer is relevant to the consortium as a whole.
3	In point 1.3 is stated: Subcontracting is allowed only up to 30% of the contract amount. → Does this option allows the economic operator/consortium to subcontract company other than the ones provided in the consortium?	Yes, it does. Note that the threshold of 30% does only apply to subcontracting not yet to the repartition of tasks amongst the consortium members.
4	Regarding your requirements in 11.2.1 Specific requirements related to the documentation accompanying the technical offer of the equipment (limited to Lot 03) do we need to submit original user and service manuals for all instruments or for some basic instrument such as microscopes, refrigerators etc.. didactic material is enough.	Requirement 11.2.1 of ITT is amended as follows: Point 2 "The awarded Supplier(s) shall provide with the delivery of the equipment complete and original user and service manuals, in hard copy and CD/DVD support, in English language and, if available, in Albanian and Serbian."
5	Regarding ITEM A03 Deep Freezer, Ultralow Temperature (UMDNS 22068)- 18) Shall have double-layer door protecting air leakage → Is it possible to have different mechanism doing the same protection of air leakage?	No, the equipment shall be compliant to the specifications.
6	. ITEM A04 Refrigerator, Laboratory, 400 lt (UMDNS 17157)- 14) Microprocessor controlled equipment, integrated data logger, → Can the data logger be external and validated instead of internal? This will make easier for yearly validation and can reduce the price while is doing the same purpose. According all ISO guidelines and standards external data looger is sufficient as long as it is validated. 16) Automatic diagnostic self-test → Is this criterion mandatory or preferable as the refrigerator has alarms, sensors and data logger that can report the conditions. With this criterion only the price of the equipment is increased?	14) external validated data logger, certified according to applicable ISO standards, will be accepted. 16) The equipment shall be capable of self-test and shall have alarms and sensors as per specifications.
7	ITEM A09 Water Purification System (UMDNS 15612)- 5) 3 Phase 400 V 50 Hz., with line connection plug IEC 309 → Is 3 phase mandatory as 1 phase is normal electricity output as it is in other equipments?	5) 3 phase power supply as per specifications is required.

8	<p>ITEM A10 Analyser, Laboratory, Immunochemistry, Automated (UMDNS 18625)-</p> <p>18) Throughput: approx 35 tests/hour → Is 30 tests/hour acceptable as approx. as it makes no significant input on the daily routine and throughput, but can significantly reduce the cost and is allowing wider range of machines?</p> <p>19) Direct sampling and sample tray accepting cups or tubes in a wide range of sizes → Can this criterion be “and/or sample tray...” instead of “and sample tray...” as various machines work in different way w/o affecting the sensitivity and specificity of the equipment and is allowing wider range of machines?</p> <p>23) Liquid level sensing → Some machines, fully compliant with the technical requirements of the ITEM 10A do not use external liquids, but instead have fully closed cartridges. As such they do not require liquid level sensing as this is used in machines that have open liquid reagents instead. Hence, we ask to put this criterion as not mandatory, but if needed for performance.</p>	<p>18) 30 tests/hour, calculated for the highest time-requiring tests in normal working conditions, is acceptable.</p> <p>19) The equipment shall have a sample tray accepting cups or tubes of various sizes, and shall be suitable for loading samples directly from the tubes/cups in the sample tray.</p> <p>23) Equipment with sealed cartridges are accepted. Point 23) is amended as follows: “Liquid level sensing, in case of device equipped with tanks for reagents; if the device is equipped with reagent cassettes, it shall be suitable for monitoring the cassette use and estimated life, and for notifying the cassette malfunction/forthcoming depletion/need for replacement.”</p>
9	<p>. ITEM A13 Microscopes, Light, Laboratory (UMDNS 15156)-</p> <p>14) The body shall be epoxy powder coated or better finished, heat treated, resistant to standard reagents used for staining, resistant to disinfecting agents for cleaning of lenses., → Better finished body is difficult to explain in papers, does this mean that the body should be covered with something specific and what would be the requirements to prove this?</p> <p>23) All objectives should be engraved with the following information: → Does this criterion mean physically engraved or it can be specially printed and labeled by the manufacturer? Engraving is significantly costly.</p>	<p>14) Epoxy powder coating, heat treated and compliant with other specifications, is required. Better finishing is intended as other finishing method which is compliant with other specifications and also compliant to applicable ISO and EC standards for medical laboratory equipment. If non epoxy coated finishing is offered, additional original documentation and certification supporting the evaluation shall be provided.</p> <p>23) All objectives shall be permanently engraved or marked (laser, physical engraving or equivalent technique) by the Manufacturer with, at least, the information provided in the specification. Marking/engraving shall be permanent, resistant to cleaning, disinfection, contact with reagents, chemical and biological products, and also to wear and accidental scratch/abrasion. Labels or any other non-permanent marking are not acceptable.</p>
10	<p>. ITEM A14 Centrifuges, Tabletop, Refrigerated (UMDNS 18265)</p> <p>Additional consumables and accessories for each unit to be included in the offer</p> <p>1) Angle rotor, compatible with standard round bottom blood tubes 12 places max capacity 12x15 ml → Can this be and/or swing rotor or only angle rotor? Swing rotors do the same job but are more flexible with adapters for various tube types as needed below.</p>	<p>1) and 2) Angle rotor with adapters is required.</p>

	2) Adapters for 2 ml, 5 ml and 7 ml	
11	Lot#3, items A10, A11 and A15 are dedicated exclusively for manufacturer ***** and except one vendor company no one can offer for those items. Is it possible to be divided in separate lot those three items?	The statement is not true, since these products are available from many Manufacturers, and there's no need to create a new lot.